



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

MESS S.C.  
 Acceso III, Número 16A Nave 10  
 Parque Industrial Benito Juárez  
 Querétaro, Qro. 76120, MÉXICO  
 Oscar Morales Phone: 52 442 476 0646, 52 442 196 4938

CALIBRATION

Valid To: May 31, 2025

Certificate Number: 3047.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1,9</sup>:

I. Dimensional

Parameter/Equipment	Range	CMC <sup>2,5</sup> (±)	Comments
Coordinate Measuring Machines (CMM) Verification <sup>3</sup> –			
E <sub>L</sub> - Length Measurement Error	Up to 2000 mm Up to 12 000 mm	(0.52 + 0.90L) μm (0.29 + 0.60L) μm	ASME B89.4.10360-2, ISO 10360-2 with step gage, gage blocks, laser interferometer
Rotary Table (4 <sup>th</sup> axis) – Four-axis Errors (FR, FT, FA)	Distance between Sphere: Up to 500 mm and CMM Resolution >= 0.1 μm	1 μm	ISO 10360-3 with test spheres
Single-Stylus Probing Size Errors	Spheres with Nominal Diameter: 25 mm	0.60 μm	ISO 10360-5 with master sphere
Single-Stylus Probing Form Errors	Spheres with Nominal Diameter: 25 mm	0.20 μm	
Scanning Mode Size Errors	Spheres with Nominal Diameter: 25 mm	0.60 μm	
Scanning Mode Form Errors	Spheres with Nominal Diameter: 25 mm	0.20 μm	

Parameter/Equipment	Range	CMC <sup>2, 5, 10</sup> ( $\pm$ )	Comments
Surface Finish (Roughness) Measuring Machines (Profilometers) <sup>3</sup>	Ra: (0.3 to 3.2) $\mu\text{m}$ Rz: (1.5 to 10) $\mu\text{m}$ Pt: (0.36 to 2.6) $\mu\text{m}$ Rsm: (15 to 100) $\mu\text{m}$	0.048 $\mu\text{m}$ 0.080 $\mu\text{m}$ 0.080 $\mu\text{m}$ 0.080 $\mu\text{m}$	ISO 12179 with surface finish standards, optical flat
Surface Geometry Analyzers (Contours Instruments) <sup>3</sup>	Angle: 90° X = Up to 1 mm X = (1 to 10) mm X = (10 to 200) mm Z = (1 to 50) mm Radius: (2.5 to 80) mm	1.5' 1.3 $\mu\text{m}$ 1.6 $\mu\text{m}$ (1.7 + 4.5L) $\mu\text{m}$ (0.12 + 5.0L) $\mu\text{m}$ (1.7 + 4.5L) $\mu\text{m}$	ISO 12179 with gage blocks, master spheres, contour standards
Roughness Standards –  Z Axis Average Parameters, Ra  Z Axis Point Parameters, Rz  Hybrid Parameters, RSm  Dimensional Material Ratio Parameters, Rk  Percentage Material Ratio Parameters	Up to 3.5 $\mu\text{m}$ >3.5 $\mu\text{m}$ to 6.5 $\mu\text{m}$  Up to 25 $\mu\text{m}$  Up to 150 $\mu\text{m}$  Up to 4 $\mu\text{m}$  0 % to 100 %	(13 + 0.003 Ra) nm (0.01 Ra - 11) nm  (35 + 0.005 Rz) nm  (60 + 0.001 RSm) nm  (10 + 0.05 Rk) nm  1 %	Taylor Hobson NOVUS PGI  <i>Ra, Rz, RSm, Rk are values in nm</i>
Gage Blocks –  Steel:  Deviation of Central Length  Variation in Length	(0.50 to 10) mm (>10 to 25) mm (>25 to 50) mm (>50 to 75) mm (>75 to 100) mm  (0.50 to 10) mm (>10 to 25) mm (>25 to 50) mm (>50 to 75) mm (>75 to 100) mm	0.024 $\mu\text{m}$ 0.030 $\mu\text{m}$ 0.044 $\mu\text{m}$ 0.062 $\mu\text{m}$ 0.081 $\mu\text{m}$  0.022 $\mu\text{m}$ 0.023 $\mu\text{m}$ 0.025 $\mu\text{m}$ 0.028 $\mu\text{m}$ 0.031 $\mu\text{m}$	NMX-CH-3650-IMNC ASME B89.1.9:  Gage blocks-ISO; Grade K, electromechanical comparator

Parameter/Equipment	Range	CMC <sup>2, 5</sup> ( $\pm$ )	Comments
Gage Blocks – (cont)			NMX-CH-3650-IMNC ASME B89.1.9:
Ceramic:			
Deviation of Central Length	(0.50 to 10) mm (>10 to 25) mm (>25 to 50) mm (>50 to 75) mm (>75 to 100) mm	0.024 $\mu$ m 0.031 $\mu$ m 0.048 $\mu$ m 0.067 $\mu$ m 0.087 $\mu$ m	Gage blocks-ISO; Grade K, electromechanical comparator
Variation in Length	(0.50 to 10) mm (>10 to 25) mm (>25 to 50) mm (>50 to 75) mm (>75 to 100) mm	0.022 $\mu$ m 0.023 $\mu$ m 0.024 $\mu$ m 0.026 $\mu$ m 0.029 $\mu$ m	
Micrometer	Up to 1000 mm	(0.85 + 16L) $\mu$ m	JIS 7502:2016, gage blocks
Calipers	Up to 1000 mm	(1.0 + 21L) $\mu$ m	ISO 13385-1, ISO 13385- 2, gage blocks with ring gage
Articulated Arm Coordinate Measuring Machines (AACMM) <sup>3</sup> –			ASME B89.4.22:
Volumetric Performance, Radius	Up to 1500 mm	(4.5 + 7.0L) $\mu$ m	Length standards, step gages
Effective Diameter Performance, Sphere Diameter	30 mm (nominal)	2.0 $\mu$ m	Master sphere

Parameter/Equipment	Range	CMC <sup>2, 5</sup> ( $\pm$ )	Comments
Roundness Measuring Machines <sup>3</sup> –			ISO 4291:
Sensitivity	(0.4 to 10) $\mu\text{m}$ (160 to 500) $\mu\text{m}$	65 nm + 46 nm/ $\mu\text{m}$ 0.090 $\mu\text{m}$	Optical flat, gage blocks
Radial Error	(6 to 25) mm	0.050 $\mu\text{m}$	Roundness standard
Axial Error	(15 to 70) mm	0.040 $\mu\text{m}$	Optical flat
Optical Comparators <sup>3</sup> – Vision Systems & Measuring Microscopes			JIS B 7184:
X and Y Axis – Error of Indication	Up to 300 mm (300 to 500) mm	(1.1 + 3.5L) $\mu\text{m}$ (3.0 + 4.5L) $\mu\text{m}$	Glass scales
Z Axis – Error of Indication	Up to 300 mm	(2.4 + 3.2L) $\mu\text{m}$	Gage blocks
Angle <sup>6</sup>	Up to 180°	1.3'	Angular reticule
Universal Length Machine (ULMs, UMMs) <sup>3</sup>			
Res. $\geq$ 0.01 $\mu\text{m}$	Up to 1000 mm	(0.25 + 2.0L) $\mu\text{m}$	Gage blocks
Res. $\geq$ 0.01 $\mu\text{m}$	Up to 2000 mm	(0.25 + 1.6L) $\mu\text{m}$	Laser interferometer
Res. $\geq$ 1 $\mu\text{m}$	Up to 5000 mm	(1.3 + 1.5L) $\mu\text{m}$	Laser interferometer
Contour Standards –			
X, Z Length Radius Angles	Up to 220 mm Up to 80 mm Up to 180°	(3.0 + 15L) $\mu\text{m}$ (3.0 + 15L) $\mu\text{m}$ 0.01°	Taylor Hobson NOVUS PGI
Roundness Standards –			
Radius	Up to 150 mm	46 nm	ISO 4291 by Error Separation Method with roundness/cylindrical profile measuring instrument

Parameter/Equipment	Range	CMC <sup>2,5</sup> ( $\pm$ )	Comments
Length Standards (Bars, Rods, Micrometers Setting Standards, and Fixtures Used as Length Standards)	(10 to 550) mm	$(0.81 + 1.0L) \mu\text{m}$	Universal length machine (ULM)
	Up to 1500 mm	$(1.7 + 7.0L) \mu\text{m}$	CMM used as comparator (substitution method), step gages
Surface Plates (Granite and Cast Iron) <sup>3</sup> –  Overall Flatness	Size of Plate: (160 x 100) mm to (4000 x 1600) mm	$(1.7 + 1.6L) \mu\text{m}$	Partial calibration per NMX-CH-8512-1:IMNC, NMX-CH-8512-2:IMNC with photoelectric autocollimator, electronic differential levels; <i>L</i> is the diagonal measurement
Height Gages	Up to 1000 mm	$(5.7 + 3.2L) \mu\text{m}$	ISO 13225:2012, gage blocks
Length Indicators – (Dial, Test/Lever, Linear Amplifiers)			Mitutoyo I-Checker:
Res. = 0.001 mm	0.1 $\mu\text{m}$ to 0.2 mm	0.35 $\mu\text{m}$	ISO 9493
Res. = 0.001 mm	(0.2 to 101.6) mm	$(1.1 + 2.0L) \mu\text{m}$	ISO 463, ISO 13102
Res. = 0.01 mm	0.1 $\mu\text{m}$ to 1.6 mm	2.6 $\mu\text{m}$	ISO 9493
Cylindrical Gages – Plain Ring Gages			ANSI/ASME B89.1.6:
Internal Diameter Class "Z, ZZ"	(3 to 150) mm	1.4 $\mu\text{m}$	Universal length machine (ULM)
Roundness:			
Diameter	Up to 400 mm	0.15 $\mu\text{m}$	Roundness machine
Height	Up to 500 mm	0.15 $\mu\text{m}$	
Cylindrical Gages – Plain External Diameters (Plug Gages, Disks, Pin Thread Wires)	(0.1 to 100) mm	$(0.48 + 5.2L) \mu\text{m}$	Universal length machine (ULM)

Parameter/Equipment	Range	CMC <sup>2, 5</sup> (±)	Comments
Spheres – Diameter	(1.5 to 100) mm	(0.41 + 5.0L) μm	ISO 3290-1, ISO 3290-2: Class G 10 to G 200, universal length machine (ULM)
Roundness: Diameter	Up to 400 mm	0.15 μm	Roundness machine
Height	Up to 500 mm	0.15 μm	
Feeler Gauges/Thickness Gauges – Plastic	(0.01 to 5) mm	(0.65 + 90L) μm	JIS B 7524; universal length machine (ULM)
Steel	(0.01 to 5) mm	(0.37 + 90L) μm	
Wire Test Sieve Cloth and Test Sieves – XY Openings	(0.075 to 40) mm	(2 + 220L) μm	NMX-CH-012-1-INMC- 2007, ASTM E11; vision system
Cylindrical Gages – Thread Plugs Simple Pitch Diameter	(0.35 to 8) mm (4 to 64) tpi (3 to 40) tpi (2 to 16) tpi	(2.5 + 3.0L) μm	Universal length machine (ULM), thread wires; metric and unified 60° threads Whitworth 55°, ACME 29°
Major Diameter	Up to 100 mm	(2.5 + 3.0L) μm	Universal length machine (ULM)
Cylindrical Gages – Thread Rings Simple Pitch Diameter	(0.6 to 3) mm (10 to 40) tpi (10 to 16) tpi	2.2 μm	Universal length machine (ULM), spherical probe Metric and unified 60° threads
Minor Diameter	(3 to 100) mm	2.2 μm	Whitworth 55°, ACME 29°

Parameter/Equipment	Range	CMC <sup>2, 5</sup> (±)	Comments
Radius Gages	0.5 µm to 25.4 mm	1.6 µm	Direct measurements with vision system
Spherical Caps – Radius Form	(5 to 95) mm (0.1 to 5.0) µm	1.5 µm 0.25 µm	Substitution method with CMM and standard sphere or standard cap
Harmonics and Cylindrical Roughness Standards – Wavelength Amplitude Ra Rz RSm	≤ 10 µm Up to 3.5µm Up to 25 µm Up to 150 µm	40 nm 20 nm 25 nm 90 nm	Taylor Hobson NOVUS PGI
Magnification or Flick Standards Calibration – Static Method Dynamic Method	(0.1 to 500) µm (0.1 to 500) µm	75 nm 250 nm	Taylor Hobson NOVUS PGI Taylor Hobson, Talyrond 595H

II. Dimensional Testing/Calibration<sup>1</sup>

Parameter/Equipment	Range	CMC <sup>2, 5</sup> (±)	Comments	
Length – 1D, 2D, 3D Measurements (Size, Position and Geometric/Form) <sup>4</sup>	Up to 2860 mm (diagonal): X = Up to 2400 mm Y = Up to 1200 mm Z = Up to 1000 mm	(2.5 + 10L) μm	CMM, hand tools, part drawings, customer requirements	
	Up to 1500 mm Up to 2700 mm Up to 3700 mm	34 μm 55 μm 91 μm	Articulating arm, CMM (AACMM), part drawings, customer requirements	
Profile and Surface Texture <sup>4</sup> –	Profile	Z: Up to 50 mm X: Up to 200 mm	2.0 μm	
	Surface Texture Up to a 30 mm Scan	Z: Amplitude/Average parameters (R, W, P <sub>a</sub> , P <sub>q</sub> )	50 nm	
		Z: Point/Spacing parameters (R <sub>Sm</sub> , W <sub>Sm</sub> , P <sub>Sm</sub> , HSC, P <sub>c</sub> )	150 nm	
		Hybrid parameters	140 nm	
Geometry/Form (Roundness, Cylindricity Straightness, Flatness & Parallelism) <sup>4</sup>	Probe Arm Range: Diameter = Up to 400 mm Height Z = Up to 500 mm	0.15 μm	Roundness machine, CMM; part drawings and customer requirements	
Non-Contact Measurements –	Angle	Angle 0° a 360°	3.5' arc (1.5 + 3.0L) μm (1.6 + 3.0L) μm	Vision system
	Length Radius	Up to 640 mm Up to 300 mm		



III. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC <sup>2,7</sup> (±)	Comments
DC Voltage – Generate	(1 to <330) mV (0.33 to <3.3) V (3.3 to <33) V (33 to <330) V (330 to 1) kV	40 μV/V + 2.0 μV 22 μV/V + 10 μV 24 μV/V + 200 μV 36 μV/V + 6.0 mV 36 μV/V + 10 mV	Fluke 5522A
DC Voltage – Measure	(10 to 100) mV (>0.1 to 1) V (>1 to 10) V (>10 to 100) V (>0.1 to 1) kV	18 μV/V + 1.0 μV 13 μV/V + 4.3 μV 14 μV/V + 96 μV 19 μV/V + 330 μV 19 μV/V + 4.6 mV	Transmille 8081 DMM
DC Current – Generate	(0.02 to 0.2) mA (>0.2 to 2) mA (>2 to 20) mA (>20 to 200) mA (>0.2 to 2) A (>2 to 20) A (>20 to 30) A	0.2 mA/A + 20 nA 0.1 mA/A + 230 nA 0.1 mA/A + 3.0 μA 0.1 mA/A + 23 μA 0.26 mA/A + 120 μA 0.6 mA/A + 600 μA 1.0 mA/A + 900 μA	Transmille 4010
Resistance – Generate	(1 to <11) Ω (11 to <110) Ω (0.11 to <1.1) kΩ (1.1 to <11) kΩ (11 to <110) kΩ (0.11 to <1.1) MΩ (1.1 to <3.3) MΩ (3.3 to <11) MΩ (11 to <33) MΩ (33 to <110) MΩ (110 to <330) MΩ	80 μΩ/Ω + 20 mΩ 60 μΩ/Ω + 30 mΩ 56 μΩ/Ω + 40 mΩ 56 μΩ/Ω + 400 mΩ 56 μΩ/Ω + 2 Ω 64 μΩ/Ω + 22 Ω 120 μΩ/Ω + 820 Ω 260 μΩ/Ω + 500 Ω 500 μΩ/Ω + 46 kΩ 1 mΩ/Ω + 100 kΩ 6 mΩ/Ω + 200 kΩ	Fluke 5522A



Parameter/Range	Frequency	CMC <sup>2,7</sup> ( $\pm$ )	Comments
AC Voltage – Generate (1 to <33) mV (33 to <330) mV (0.33 to <3.3) V (3.3 to <33) V (33 to <330) V (0.33 to 1) kV  (33 to <330) mV (0.33 to <3.3) V (3.3 to <33) V (33 to <330) V	45 Hz to 10 kHz  (>10 to 20) kHz	0.30 mV/V + 12 $\mu$ V 0.30 mV/V + 94 $\mu$ V 0.30 mV/V + 920 $\mu$ V 0.30 mV/V + 11 mV 0.38 mV/V + 90 mV 0.60 mV/V + 56 mV  0.32 mV/V + 30 $\mu$ V 0.38 mV/V + 230 $\mu$ V 0.48 mV/V + 3.8 mV 0.50 mV/V + 12 mV	Fluke 5522A
AC Voltage – Measure (10 to 100) mV (>0.1 to 1) V (>1 to 10) V (>10 to 100) V (>0.1 to 1) kV	40 Hz to 1 kHz	0.6 mV/V + 18 $\mu$ V 0.6 mV/V + 120 $\mu$ V 0.6 mV/V + 1.2 mV 0.6 mV/V + 18 mV 0.6 mV/V + 180 mV	Transmille 8081 DMM
AC Current – Generate (0.02 to 0.2) mA (>0.2 to 2) mA (>2 to 20) mA (>20 to 200) mA (>0.2 to 2) A (>2 to 30) A	45 Hz to 1 kHz	1.4 mA/A + 1.1 $\mu$ A 1.2 mA/A + 1.7 $\mu$ A 0.8 mA/A + 4 $\mu$ A 0.8 mA/A + 240 $\mu$ A 1.2 mA/A + 4.8 mA 1.6 mA/A + 16 mA	Transmille 4010
AC Current – Measure (0.1 to 1) mA (>1 to 10) mA (>10 to 100) mA (>0.1 to 1) A (>1 to 10) A (>10 to 30) A	40 Hz to 1 kHz	1.0 mA/A + 240 nA 1.0 mA/A + 2.4 $\mu$ A 1.0 mA/A + 24 $\mu$ A 1.4 mA/A + 300 $\mu$ A 2.4 mA/A + 6 mA 2.4 mA/A + 18 mA	Transmille 8081 DMM

Parameter/Equipment	Range	CMC <sup>2,7</sup> ( $\pm$ )	Comments
DC Current – Measure	(0.1 to 1) $\mu$ A ( $>1$ to 10) $\mu$ A ( $>10$ to 100) $\mu$ A ( $>0.1$ to 1) mA ( $>1$ to 10) mA ( $>10$ to 100) mA ( $>0.1$ to 1) A ( $>1$ to 10) A ( $>10$ to 30) A	1.4 mA/A + 68 pA 200 $\mu$ A/A + 400 pA 56 $\mu$ A/A + 5.8 nA 56 $\mu$ A/A + 62 nA 64 $\mu$ A/A + 660 nA 94 $\mu$ A/A + 1.2 $\mu$ A 470 $\mu$ A/A + 26 $\mu$ A 1.1 mA/A + 700 $\mu$ A 1.5 mA/A + 9 mA	Transmille 8081 DMM
Resistance – Measure	(0.1 to 1) $\Omega$ ( $>1$ to 10) $\Omega$ ( $>10$ to 100) $\Omega$ ( $>0.1$ to 1) k $\Omega$ (1 to 10) k $\Omega$ ( $>10$ to 100) k $\Omega$ ( $>0.1$ to 1) M $\Omega$ ( $>1$ to 10) M $\Omega$	94 $\mu\Omega/\Omega$ + 24 $\mu\Omega$ 63 $\mu\Omega/\Omega$ + 870 $\mu\Omega$ 56 $\mu\Omega/\Omega$ + 7.9 m $\Omega$ 25 $\mu\Omega/\Omega$ + 20 m $\Omega$ 30 $\mu\Omega/\Omega$ + 400 m $\Omega$ 31 $\mu\Omega/\Omega$ + 1.6 $\Omega$ 36 $\mu\Omega/\Omega$ + 56 $\Omega$ 48 $\mu\Omega/\Omega$ + 880 $\Omega$	Transmille 8081 DMM
Electrical Simulation of Thermocouple Indicators –  Type B Type C Type E Type J Type K Type N Type R Type S Type T	(600 to 1820) $^{\circ}$ C (0 to 2316) $^{\circ}$ C (-250 to 1000) $^{\circ}$ C (-210 to 1200) $^{\circ}$ C (-200 to 1372) $^{\circ}$ C (-200 to 1300) $^{\circ}$ C (0 to 1767) $^{\circ}$ C (0 to 1767) $^{\circ}$ C (-250 to 400) $^{\circ}$ C	0.18 $^{\circ}$ C 0.16 $^{\circ}$ C 0.094 $^{\circ}$ C 0.094 $^{\circ}$ C 0.10 $^{\circ}$ C 0.11 $^{\circ}$ C 0.20 $^{\circ}$ C 0.21 $^{\circ}$ C 0.094 $^{\circ}$ C	Fluke 5522A

IV. Mechanical

Parameter/Equipment	Range	CMC <sup>2,8</sup> (±)	Comments
Force Measuring Devices <sup>3</sup> –  Tension and Compression	(0.2 to 25) N (>25 to 500) N (>500 to 1600) N  (1 to 10) kN (>10 to 100) kN (33 to 334) kN (89 to 890) kN	0.032 % of rdg 0.0039 % of rdg 0.0026 % of rdg  0.026 % of rdg 0.020 % of rdg 0.053 % of rdg 0.065 % of rdg	ISO 376:2004, NMX-CH-376-I-IMNC-2008:  Dead weights used as mass standards  Load cells and digital indicators
Force Testing Machines –  Tension and Compression  Compression Only	(0.2 to 25) N (>25 to 500) N (>500 to 1600) N  (1 to 10) kN (>10 to 100) kN (33 to 334) kN (89 to 890) kN  (222 to 2224) kN	0.032 % of rdg 0.0039 % of rdg 0.0026 % of rdg  0.026 % of rdg 0.020 % of rdg 0.053 % of rdg 0.065 % of rdg  0.091 % of rdg	ISO 7500-1:2004, NMX-CH-7500-1-IMNC-2008:  Dead weights used as mass standards  Load cells and digital indicators  Load cells and digital indicators
Torque Instruments –  Hand Torque Tools, Torque Wrenches	(0.2 to 2) N·m (1 to 10) N·m (2.5 to 25) N·m (15 to 150) N·m (40 to 400) N·m (150 to 1500) N·m	(0.33 to 0.19) % of rdg (0.27 to 0.18) % of rdg (0.32 to 0.16) % of rdg (0.29 to 0.15) % of rdg (0.23 to 0.15) % of rdg (0.33 to 0.16) % of rdg	ISO 6789:2003; NMX-CH-6789-IMNC-2006:  Torque transducers with indicators  <i>Uncertainties at intermediate values by linear interpolation.</i>
Torque – Measuring Equipment (Torque Transducers/Torque Analysers)	(50 to 150) N·m (150 to 500) N·m (500 to 2400) N·m	(0.17 to 0.14) % of rdg (0.14 to 0.04) % of rdg (0.087 to 0.029) % of rdg	Mass as dead weights, loading arm  <i>Uncertainties at intermediate values by linear interpolation.</i>

## V. Thermodynamic

Parameter/Equipment	Range	CMC <sup>2, 8</sup> (±)	Comments
Temperature – Measuring Equipment	(-45 to 0) °C	0.060 °C	Fluke 5626 SPRT with Fluke 8845A DMM and Fluke 9170-9173 dry blocks
	(>0 to 140) °C	0.080 °C	
	(>140 to 660) °C	0.10 °C	
	(>660 to 700) °C	0.25 °C	
			Fluke 9173 Dry block
Infrared Thermometers	(35 to 100) °C	0.45 °C	Fluke 4181 IR calibrator $\epsilon = 0.95$ , $\lambda = (8 \text{ to } 14)$ $\mu\text{m}$
	(>100 to 200) °C	0.64 °C	
	(>200 to 350) °C	1.1 °C	
	(>350 to 500) °C	1.5 °C	

## VI. Time & Frequency

Parameter/Equipment	Range	CMC <sup>2, 8</sup> (±)	Comments
Frequency – Generate  (1 to 5) Vp-p	(10 to 120) Hz	0.29 Hz	Fluke 5522A
	(120 to 1000) Hz	3.5 Hz	
	(1 to 10) kHz	0.029 kHz	

## VII. Mechanical Testing

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following test on metals, plastic:

### Test Description

### Test Method

Tension and Compression Forces

ISO 6892-1, ISO 7438; MESS-FZ-PRO-003

Tensile Yield Strength, Elongation

ISO 6892-1, ISO 7438; MESS-FZ-PRO-003

- <sup>1</sup> This laboratory offers commercial dimensional testing/calibration service and field calibration service.
- <sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.
- <sup>3</sup> Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g., resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.
- <sup>4</sup> This laboratory meets R205 – *Specific Requirements: Calibration Laboratory Accreditation Program* for the types of dimensional tests listed above and is considered equivalent to that of a calibration.
- <sup>5</sup> In the statement of CMC,  $L$  is the numerical value of the nominal length of the device measured in meters.
- <sup>6</sup> Applicable to optical comparators.
- <sup>7</sup> The stated measured values are determined using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMCs are expressed as either a specific value that covers the full range or as a percent or fraction of the reading plus a fixed floor specification.
- <sup>8</sup> The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.
- <sup>9</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.
- <sup>10</sup> In the statement of CMC, percentages are to be read as percent of indicated value.



# Accredited Laboratory

A2LA has accredited

**MESS S.C.**

*Queretaro, MEXICO*

for technical competence in the field of

**Calibration**

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 19<sup>th</sup> day of July 2023.

A blue ink signature of Mr. Trace McInturff, written in a cursive style.

Mr. Trace McInturff, Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 3047.01  
Valid to May 31, 2025

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*